



# SMJ


## Siriraj Medical Journal

The world-leading biomedical science of Thailand


MONTHLY

ORIGINAL ARTICLE

### Comparing Reusable and Single-use ureteroscope



#### Comparing Mechanical and Optical characteristics



#### Results

	XC1	XC
Deflection	<	
Loop diameter	=	
Irrigation flow rate	<	
Resolution	=	
Color	>	

#### Optical characteristics



#### Mechanical characteristics



Conclusion: The XC1 offered improved color representation and marginally better resolution in saline.

SCAN FOR  
FULL TEXT



SIRIRAJ  
MEDICAL  
JOURNAL | Leelathanyakit, et al. *Siriraj Med J* 2025;77(5):313-321.

©Siriraj Medical Journal. All rights reserved. Graphical abstract by K Suriyapornpun

Indexed by

Scopus®

DOAJ



<https://he02.tci-thaijo.org/index.php/sirirajmedj/index>  
E-mail: [sijournal92@gmail.com](mailto:sijournal92@gmail.com)